



Product Service

Attestation of Conformity

No. N8A 16 09 83777 012

Holder of Certificate: **Associated Research, Inc.**

13860 West Laurel Drive
 Lake Forest IL 60045
 USA

Product: **Laboratory Equipment
 (Ground Bond Tester)**

Model(s): **HYAMP series 3240**

Parameters:

Rated voltage:	100-120/200-240 Vac
Rated frequency:	50/60 Hz
Rated power:	Max. 850 VA
Rated output:	1.0-10.0A/0-600m ohm, 10.1-30.0A/0-200m ohm, 30.1-40.0A/0-150m ohm
Protection class:	I

Tested according to: EN 61010-1:2010
 EN 61010-2-030:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 611001601301

Bill Lin

Date, 2016-09-06 (Bill Lin)



CE After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.